



Republic of the Philippines
Benguet State University
2601 La Trinidad, Benguet
www.bsu.edu.ph

Telefax No. (074) 661-1839; Email:procurement@bsu.edu.ph

REQUEST FOR QUOTATION (RFQ)
(ALL ENTRIES ARE MANDATORY)

Name of Owner/ Corporation/ Cooperative/ Agency: _____ QUOTATION NO. 1432
Name of Business : _____ DATE: SEP 02 2026
Address: _____ P.R. NO. : 2026-07-IGF-574 ✓
Email address: _____ ABC: Php. 220,000.00
Telephone No. _____ FUND SOURCE: IGI

The Benguet State University through its Bids and Awards Committee will undertake SVP
_____ for the item/s listed below under the following terms and conditions:

1. All entries must be written legibly and accurately.
2. Delivery period is within 60 calendar days.
3. Warranty period shall be for a minimum of three months, in case of expendable supplies, and a minimum of one year in case of non-expendable supplies from date of acceptance by the University.
4. Price validity shall be for a period of 45 calendar days.
5. Compliance with specifications: Kindly specify in the column provided below whether the offer fully complies with the specifications or a counter offer.
6. Deadline of submission of quotation is on or before 5:00 PM of Sept. 17, 2026. Offers may be submitted in sealed or open quotation in person, or through facsimile (074-661-1839) or email (procurement@bsu.edu.ph)
7. The item/s shall be procured by: _____ a) line item; / _____ b) lot; _____ c) sub-lot _____, place of issue: _____
8. Business Permit Number : _____, date of issue: _____, date of renewal: _____, place of issue: _____
(If renewal of Business Permit is still on process, please indicate OR No: _____, date of OR : _____ for the payment of renewal).
9. * PhilGEPS Registration Number: _____

College/Dept./Office: **CVM - Veterinary Teaching Hospital**

JOSE M. FLORENTIN
Chairperson, BAC (Goods & Services)

Item No.	Qty	Unit	Item	Technical Specifications (Brand and Model)		ABC	Unit Cost	Total Cost
				COMPLY	COUNTER OFFER			
	1	unit	Veterinary electric turbo ventilator Multiple breathing modes: VCV, PCV, SIMV, SPONT; Built in digital breathing sensor for accurate monitoring; with 1/2" santadrad bellows to meet ventilation needs of different animals; 8-inch HD touch screen for easy human-machine interaction; Three levels of color and light alarm; Database of standards for animals, can connect to high pressure gas source, can attach to different anaesthesia machines. Tidal volume (rv) -20-1500 ml; Breath rate -2-60BPM; Inspiratory time:Inspiratory time -4:1-1:8; Positive end expiratory pressure-OFF/3-30cm H2O; Inspiratory pressure (Pinsp) -5-50 mH2O; Inspiratory Plateau time <30s; inspiratory Time -1.2-10s; Pressure Limit Level (PLimit) -10-60cmH2O; Flow Trigger (Trig) -OFF/1-20L/min; Apnea time (Apn.t) -15-30s. Inclusive of 1 year warranty on parts and services, UPS and training on the use of the equipment			175,000.00		
2	1	unit	Ultrasound diagnostic Cardiac Table Description Frame material: steel with a polymer-powder coating Tabletop material: PVC Dimensions (table on the wheels): 132 x72 cm Tabletop size: 132x72 cm Height adjustment: 33-113 cm			45,000.00		
*			Indicate Brand & model of the offer TOTAL			220,000.00		
			XXXXXXXXXXXX					

Counter offers:

Delivery Period: _____
Warranty: _____
Price Validity: _____

Signature over Printed Name

Date and Time of Receipt of RFQ

* Not a ground for disqualification for failure to indicate but required as a condition for award.